

1/CPR/2023/z1

Katowice, 20/12/2023

Modification of the request for quotation

Refers to: REQUEST FOR QUOTATION No. 1/CPR/2023 of November 27, 2023. the subject of which is:

Delivery and installation of cyclotron with automatic solid state target system and synthesis modules set.

implemented as part of the project entitled: „RADIOFARM 4.0: Construction of a new intelligent radiopharmaceuticals factory” for which the Contracting Authority is applying for funding as part of an application for grant related to robotization, artificial intelligence or digitization of processes, technologies, products or services, implemented as part of the investment A2.1.1 Investments supporting robotization and digitization in enterprises from the National Recovery and Resilience Plan (KPO).

Acting on behalf of Alteris S.A., we would like to inform about modification of the content of the REQUEST FOR QUOTATION No. 1/CPR/2023 by changing a part of the content in section XII. IMPORTANT TERMS OF THE CONTRACT, CONDITIONS FOR AMENDING THE CONTRACT AND CONDITIONS FOR TERMINATION OF THE CONTRACT / WITHDRAWAL, point 6, letter a. as follows:

Was:

1. „Contractual warranty terms:
 - a. The Contractor will provide the Contracting Authority with at least twelve-month contractual warranty for the subject of the contract, counted from the moment of signing the handover protocol. During the contractual warranty period, the Contractor is obliged to repair and remove any defects or faults in the products (including replacements) revealed during the warranty period at his own expense in accordance with the notification issued by the Contracting Authority. The contractor covers the costs of any transport.

Is:

- „6. Contractual warranty terms:
 - a. The Contractor will provide the Contracting Authority with at least twelve-month contractual warranty for the subject of the contract, counted from the moment of **commencement of production activity, understood as the production of a radiopharmaceutical for sale or administration to a patient, but no later than 12 months from the moment of signing the handover protocol of the subject of the contract. The production activity does not include the following elements of acceptance testing, qualification and validation:**
 - **operational tests of the cyclotron,**
 - **hot tests required for National Atomic Energy Agency inspections in order to obtain National Atomic Energy Agency permits,**
 - **synthesizer qualifications,**

- qualifications of dispensers,
- qualifications of QC devices requiring radioactive substances,
- validation of analytical methods requiring radioactive substances,
- process validation.

During the contractual warranty period, the Contractor is obliged to repair and remove any defects or faults in the products (including replacements) revealed during the warranty period at his own expense in accordance with the notification issued by the Contracting Authority. The contractor covers the costs of any transport.”

Due to the above, the Contracting Authority posts the modified content of the Request for Quotation in Polish and English version.